
Midmark M11 - v1.0.0
Total Cycles: 2384

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2022 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.4 F 0.0
2:00 196.6 F 1.9
4:00 209.4 F 5.9
6:00 223.3 F 11.4
8:00 227.3 F 6.3
10:00 238.5 F 11.8
12:00 236.0 F 9.0
14:00 240.0 F 9.5
16:00 247.3 F 13.8
18:00 259.6 F 20.5
20:00 270.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.2 F 28.7
2:00 272.8 F 29.2
3:00 271.9 F 28.7
4:00 272.3 F 28.8

Min 271.7 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 220.8 F 0.0
5:00 225.3 F 0.0
10:00 213.5 F 0.0
15:00 229.5 F 0.0
20:00 231.7 F 0.0
25:00 230.0 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

10/03/2022 09:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____