
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002950
15:21 12/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:43
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 226kPa 4:09
136.4 C 221kPa 4:39
136.4 C 225kPa 5:09
136.2 C 222kPa 5:39
136.3 C 221kPa 6:09
136.2 C 223kPa 6:39
136.0 C 221kPa 7:09
136.0 C 221kPa 7:09
VENTING 7:09
AIR DRYING 8:00
CYCLE COMPLETE 8:02

Digital Signature #
73E305D69A64F08C
