

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3462

Name: RITTER 1  
Sterilizer ID: V2178198  
06/29/2023 07:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.1 F 0.0  
2:00 75.7 F 0.0  
4:00 81.2 F 0.0  
6:00 112.5 F 0.4  
8:00 142.7 F 1.8  
10:00 167.9 F 3.8  
12:00 185.0 F 6.5  
14:00 201.9 F 9.7  
16:00 214.7 F 13.3  
18:00 224.8 F 6.4  
20:00 233.6 F 10.6  
22:00 233.6 F 7.9  
24:00 244.1 F 12.7  
26:00 241.7 F 11.0  
28:00 254.2 F 17.4  
30:00 265.2 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.4 F 29.0  
2:00 272.4 F 28.6  
3:00 271.4 F 28.0  
4:00 271.9 F 28.6  
  
Min 271.2 F 28.0  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.7 F 0.0  
5:00 205.0 F 0.0  
10:00 186.8 F 0.0  
15:00 191.9 F 0.0  
20:00 232.8 F 0.0  
25:00 233.7 F 0.0  
30:00 230.6 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:58  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:09:56  
CYCLE COMPLETE

06/29/2023 08:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_