

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001394  
6:06 09/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:52  
PRESSURIZING 5:24  
STERILIZING 6:41  
136.0 C 220kPa 6:41  
136.5 C 222kPa 7:12  
135.9 C 218kPa 7:42  
136.4 C 218kPa 8:12  
136.1 C 219kPa 8:42  
136.4 C 222kPa 9:12  
135.8 C 217kPa 9:42  
136.4 C 221kPa 10:11  
VENTING 10:11  
AIR DRYING 11:10  
CYCLE COMPLETE 14:24

-----  
Digital Signature #  
97E4622E91B21C06  
-----