
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000457
8:49 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:11
STERILIZING 4:17
136.5 C 224kPa 4:17
136.2 C 223kPa 4:47
136.2 C 223kPa 5:17
136.2 C 223kPa 5:47
136.2 C 221kPa 6:17
136.4 C 224kPa 6:47
136.1 C 220kPa 7:17
136.3 C 223kPa 7:47
136.3 C 223kPa 7:47
VENTING 7:47
AIR DRYING 8:40
CYCLE COMPLETE 8:45

Digital Signature #
D960440D90A6988C
