
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001466
12:33 20/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:35
STERILIZING 3:31
136.0 C 220kPa 3:31
136.4 C 227kPa 4:01
136.4 C 226kPa 4:31
136.3 C 225kPa 5:01
136.4 C 226kPa 5:31
136.3 C 225kPa 6:01
136.3 C 225kPa 6:31
136.4 C 226kPa 7:01
136.4 C 226kPa 7:01
VENTING 7:01
AIR DRYING 7:52
CYCLE COMPLETE 7:57

Digital Signature #
449DE5E65458572C
