
Midmark M11 - v1.0.5
Total Cycles: 4347

Name: RITTER2
Sterilizer ID: V2313228
02/10/2025 04:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 163.5 F 2.6
4:00 189.8 F 7.1
6:00 206.4 F 11.9
8:00 220.5 F 6.4
10:00 234.1 F 11.3
12:00 233.8 F 8.4
14:00 242.5 F 11.2
16:00 245.7 F 13.5
18:00 257.7 F 20.0
20:00 268.6 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.6
2:00 271.6 F 29.0
3:00 272.5 F 29.7
4:00 271.8 F 29.0

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 230.9 F 0.0
10:00 218.0 F 0.0
15:00 217.8 F 0.0
20:00 224.9 F 0.0
25:00 215.4 F 0.0
30:00 203.5 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

02/10/2025 05:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____