

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4604

Name: RITTER 1  
Sterilizer ID: V2178198  
06/04/2024 08:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 70.9 F 0.2  
4:00 81.0 F 0.3  
6:00 110.1 F 0.5  
8:00 140.6 F 2.0  
10:00 171.1 F 5.0  
12:00 180.3 F 8.5  
14:00 197.8 F 12.0  
16:00 215.5 F 4.1  
18:00 226.0 F 7.6  
20:00 234.8 F 11.3  
22:00 230.4 F 6.8  
24:00 243.0 F 12.1  
26:00 238.9 F 9.8  
28:00 251.0 F 15.6  
30:00 261.4 F 21.5  
32:00 271.1 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 271.9 F 28.3  
2:00 272.2 F 28.8  
3:00 271.4 F 28.0  
4:00 272.0 F 28.6  
  
Min 271.2 F 28.0  
Max 272.4 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.2 F 0.0  
5:00 213.2 F 0.0  
10:00 228.6 F 0.0  
15:00 236.1 F 0.0  
20:00 224.4 F 0.0  
25:00 232.8 F 0.0  
30:00 233.8 F 0.0  
  
Drying Complete

FILLING: 0:47  
HEATING: 32:41  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:10:55  
CYCLE COMPLETE

06/04/2024 09:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_