
Midmark M11 - v1.0.5
Total Cycles: 2563

Name: RITTER2
Sterilizer ID: V2313228
06/26/2023 07:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 75.7 F 0.0
4:00 106.8 F 0.0
6:00 101.3 F 0.0
8:00 143.8 F 0.0
10:00 171.0 F 2.9
12:00 187.7 F 6.0
14:00 207.0 F 9.5
16:00 221.7 F 13.0
18:00 222.5 F 7.3
20:00 234.0 F 11.7
22:00 231.6 F 7.8
24:00 243.2 F 13.0
26:00 241.4 F 11.5
28:00 253.4 F 17.6
30:00 263.8 F 23.8
32:00 272.4 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.6
2:00 271.3 F 29.0
3:00 272.7 F 29.7
4:00 271.7 F 28.9

Min 271.3 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 208.9 F 0.0
10:00 207.8 F 0.0
15:00 207.8 F 0.0
20:00 205.8 F 0.0
25:00 209.5 F 0.0
30:00 211.0 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:13
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

06/26/2023 08:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____