
Midmark M11 - v1.0.5
Total Cycles: 1998

Name: RITTER2
Sterilizer ID: V2313228
02/02/2023 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 188.3 F 0.2
4:00 196.6 F 4.7
6:00 214.2 F 10.1
8:00 224.4 F 6.5
10:00 234.7 F 12.1
12:00 236.4 F 9.9
14:00 235.6 F 8.6
16:00 249.6 F 15.6
18:00 261.7 F 22.5
20:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.3 F 29.7
2:00 271.2 F 29.2
3:00 272.1 F 29.7
4:00 271.5 F 29.1

Min 271.2 F 29.0
Max 272.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 219.1 F 0.0
10:00 228.1 F 0.0
15:00 228.5 F 0.0
20:00 232.8 F 0.0
25:00 231.0 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

02/02/2023 11:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____