
Midmark M11 - v1.0.5
Total Cycles: 3123

Name: RITTER2
Sterilizer ID: V2313228
11/14/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.7 F 0.0
2:00 132.0 F 0.9
4:00 149.1 F 3.6
6:00 175.5 F 7.8
8:00 199.9 F 12.6
10:00 211.9 F 6.9
12:00 229.5 F 11.9
14:00 231.2 F 8.4
16:00 240.5 F 11.1
18:00 244.7 F 13.0
20:00 257.5 F 19.8
22:00 268.8 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.5
2:00 271.2 F 28.8
3:00 272.6 F 29.6
4:00 271.6 F 28.9

Min 271.1 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 212.9 F 0.0
10:00 198.2 F 0.0
15:00 194.7 F 0.0
20:00 193.5 F 0.0
25:00 192.3 F 0.0
30:00 189.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:08
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

11/14/2023 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____