
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 007687
14:18 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:37
STERILIZING 4:41
136.0 C 220kPa 4:41
136.3 C 223kPa 5:11
136.4 C 224kPa 5:41
136.2 C 223kPa 6:11
136.4 C 222kPa 6:41
136.1 C 221kPa 7:11
136.3 C 223kPa 7:41
136.3 C 222kPa 8:11
136.3 C 222kPa 8:11
VENTING 8:11
AIR DRYING 9:09
CYCLE COMPLETE 9:41

Digital Signature #
49A4ED2D30305A5D
