

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000729  
11:26 22/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:38  
STERILIZING 3:41  
136.5 C 233kPa 3:41  
136.3 C 232kPa 4:11  
136.2 C 231kPa 4:41  
136.4 C 233kPa 5:11  
136.4 C 233kPa 5:41  
136.4 C 233kPa 6:11  
136.4 C 233kPa 6:41  
136.4 C 230kPa 7:11  
136.4 C 230kPa 7:11  
VENTING 7:11  
AIR DRYING 8:03  
CYCLE COMPLETE 8:05

-----  
Digital Signature #  
E25C09A86CE5DB28  
-----