
Midmark M11 - v1.0.5
Total Cycles: 3926

Name: RITTER2
Sterilizer ID: V2313228
09/10/2024 04:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 160.4 F 2.4
4:00 182.4 F 6.3
6:00 207.6 F 11.4
8:00 219.8 F 6.3
10:00 232.6 F 11.5
12:00 233.4 F 8.9
14:00 239.6 F 10.4
16:00 246.7 F 14.0
18:00 258.9 F 20.7
20:00 269.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.3 F 29.5
2:00 271.1 F 29.2
3:00 272.0 F 29.6
4:00 272.6 F 29.7

Min 271.0 F 28.9
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 232.4 F 0.0
10:00 212.6 F 0.0
15:00 208.1 F 0.0
20:00 208.6 F 0.0
25:00 204.9 F 0.0
30:00 211.9 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

09/10/2024 05:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____