
Midmark M11 - v1.0.5
Total Cycles: 1818

Name: RITTER2
Sterilizer ID: V2313228
12/07/2022 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.3 F 0.0
2:00 168.8 F 3.9
4:00 185.0 F 8.2
6:00 208.7 F 12.1
8:00 219.8 F 8.1
10:00 233.9 F 12.2
12:00 238.2 F 11.4
14:00 238.6 F 10.3
16:00 253.1 F 17.3
18:00 264.9 F 24.6
20:00 271.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.8 F 30.2
2:00 272.2 F 29.5
3:00 272.2 F 30.0
4:00 272.8 F 30.1

Min 271.7 F 29.3
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 222.2 F 0.0
10:00 232.8 F 0.0
15:00 227.9 F 0.0
20:00 229.3 F 0.0
25:00 227.9 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

12/07/2022 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____