
Midmark M11 - v1.0.5
Total Cycles: 4150

Name: RITTER2
Sterilizer ID: V2313228
11/21/2024 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.2 F 0.0
2:00 128.4 F 0.0
4:00 152.5 F 0.8
6:00 185.1 F 4.3
8:00 203.6 F 8.0
10:00 211.1 F 12.3
12:00 224.0 F 6.8
14:00 234.5 F 11.3
16:00 233.8 F 8.4
18:00 241.6 F 11.7
20:00 244.7 F 13.2
22:00 256.4 F 19.4
24:00 267.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.8
2:00 271.4 F 29.1
3:00 272.4 F 29.9
4:00 271.7 F 29.3

Min 271.3 F 29.0
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 219.1 F 0.0
10:00 218.1 F 0.0
15:00 219.9 F 0.0
20:00 216.3 F 0.0
25:00 220.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:32
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:11
CYCLE COMPLETE

11/21/2024 05:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____