
Midmark M11 - v1.0.5
Total Cycles: 4284

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 09:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.3 F 0.0
2:00 156.9 F 2.7
4:00 183.8 F 7.1
6:00 205.7 F 12.8
8:00 224.6 F 6.8
10:00 238.7 F 12.0
12:00 236.8 F 9.5
14:00 237.9 F 9.2
16:00 250.0 F 15.3
18:00 262.4 F 22.4
20:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.3 F 28.7
2:00 272.8 F 29.2
3:00 271.8 F 28.8
4:00 272.4 F 28.9

Min 271.6 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 220.3 F 0.0
10:00 197.1 F 0.0
15:00 196.6 F 0.0
20:00 194.6 F 0.0
25:00 192.7 F 0.0
30:00 190.4 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

01/22/2025 10:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____