
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000577
7:57 12/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:37
STERILIZING 5:15
136.4 C 230kPa 5:15
136.3 C 230kPa 5:45
136.5 C 232kPa 6:15
136.2 C 231kPa 6:45
136.4 C 232kPa 7:15
136.4 C 231kPa 7:45
136.4 C 232kPa 8:15
136.5 C 232kPa 8:45
136.5 C 232kPa 8:45
VENTING 8:45
AIR DRYING 9:39
CYCLE COMPLETE 24:39

Digital Signature #
D23117BF70B62F11
