

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2625

Name: RITTER 1  
Sterilizer ID: V2178198  
11/30/2022 08:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 80.3 F 0.0  
2:00 82.7 F 0.0  
4:00 98.5 F 0.0  
6:00 122.6 F 0.1  
8:00 151.2 F 2.6  
10:00 175.4 F 6.2  
12:00 195.0 F 10.2  
14:00 210.1 F 7.8  
16:00 220.2 F 6.7  
18:00 231.5 F 10.8  
20:00 231.8 F 7.4  
22:00 242.6 F 12.5  
24:00 240.4 F 10.7  
26:00 253.2 F 17.1  
28:00 264.4 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 29.1  
2:00 271.5 F 28.5  
3:00 272.6 F 29.1  
4:00 271.6 F 28.4

Min 271.5 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.8 F 0.0  
5:00 206.7 F 0.0  
10:00 220.3 F 0.0  
15:00 224.6 F 0.0  
20:00 222.7 F 0.0  
25:00 223.4 F 0.0  
30:00 221.3 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 29:58  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:07:55  
CYCLE COMPLETE

11/30/2022 09:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_