
Midmark M11 - v1.0.5
Total Cycles: 3023

Name: RITTER2
Sterilizer ID: V2313228
10/18/2023 07:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 72.8 F 0.0
4:00 76.9 F 0.3
6:00 103.7 F 2.1
8:00 141.8 F 5.0
10:00 168.8 F 8.1
12:00 190.2 F 11.8
14:00 207.0 F 4.4
16:00 216.7 F 8.7
18:00 229.3 F 13.0
20:00 230.9 F 9.0
22:00 238.6 F 10.7
24:00 243.4 F 12.7
26:00 255.5 F 19.0
28:00 266.0 F 25.6
30:00 272.8 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.9
1:00 272.7 F 30.0
2:00 272.8 F 30.1
3:00 272.7 F 30.2
4:00 272.8 F 30.3

Min 271.9 F 29.7
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 205.8 F 0.0
10:00 212.4 F 0.0
15:00 193.2 F 0.0
20:00 195.5 F 0.0
25:00 200.4 F 0.0
30:00 203.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:13
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:11
CYCLE COMPLETE

10/18/2023 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____