
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005425
7:49 25/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:06
PRESSURIZING 4:06
STERILIZING 4:54
136.0 C 220kPa 4:54
136.4 C 223kPa 5:24
136.3 C 223kPa 5:54
135.9 C 221kPa 6:24
136.4 C 219kPa 6:54
136.4 C 223kPa 7:24
135.9 C 219kPa 7:54
136.4 C 224kPa 8:24
136.4 C 224kPa 8:24
VENTING 8:24
AIR DRYING 9:29
CYCLE COMPLETE 9:44

Digital Signature #
C07A1EFE253E244E
