
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004563
14:23 18/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:02
STERILIZING 2:55
136.0 C 220kPa 2:55
136.3 C 225kPa 3:25
136.4 C 225kPa 3:55
136.4 C 225kPa 4:25
136.1 C 221kPa 4:55
136.1 C 219kPa 5:25
136.4 C 225kPa 5:55
136.4 C 220kPa 6:25
136.4 C 220kPa 6:25
VENTING 6:25
AIR DRYING 7:15
CYCLE COMPLETE 7:20

Digital Signature #
62253F15CC32A966
