
Midmark M11 - v1.0.5
Total Cycles: 122

Name: RITTER2
Sterilizer ID: V2313228
06/21/2021 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.0 F 0.0
2:00 180.0 F 3.4
4:00 198.6 F 8.8
6:00 216.9 F 8.7
8:00 229.3 F 9.9
10:00 233.8 F 8.9
12:00 241.2 F 11.7
14:00 246.5 F 14.0
16:00 258.9 F 20.8
18:00 269.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.7 F 29.7
2:00 271.5 F 29.1
3:00 272.6 F 29.9
4:00 272.0 F 29.3

Min 271.5 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 227.7 F 0.0
10:00 230.6 F 0.0
15:00 226.9 F 0.0
20:00 225.3 F 0.0
25:00 225.7 F 0.0
30:00 209.0 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 19:00
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

06/21/2021 10:58 AM
ACCEPTED: X
REJECTED:
Operator: U01