

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000651  
11:40 19/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:23  
PRESSURIZING 3:24  
STERILIZING 4:31  
136.4 C 225kPa 4:31  
136.2 C 223kPa 5:01  
136.2 C 220kPa 5:31  
136.2 C 223kPa 6:01  
136.4 C 225kPa 6:31  
136.2 C 223kPa 7:01  
136.4 C 223kPa 7:31  
136.3 C 222kPa 8:01  
136.3 C 222kPa 8:01  
VENTING 8:01  
AIR DRYING 8:58  
CYCLE COMPLETE 9:02

-----  
Digital Signature #  
D6F0BCCA0ED2E1BA  
-----