
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000850
13:34 07/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:07
STERILIZING 3:14
136.0 C 220kPa 3:14
136.3 C 231kPa 3:44
136.3 C 232kPa 4:14
136.5 C 231kPa 4:44
136.3 C 229kPa 5:14
136.2 C 231kPa 5:44
136.2 C 230kPa 6:14
136.3 C 232kPa 6:44
136.3 C 232kPa 6:44
VENTING 6:44
AIR DRYING 7:35
CYCLE COMPLETE 22:35

Digital Signature #
0A692209AFC66A80
