
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001811
7:30 08/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:51
136.4 C 223kPa 3:51
135.9 C 220kPa 4:21
136.0 C 220kPa 4:51
136.1 C 223kPa 5:21
136.4 C 224kPa 5:51
136.2 C 221kPa 6:21
136.3 C 223kPa 6:51
136.3 C 223kPa 7:21
136.3 C 223kPa 7:21
VENTING 7:21
AIR DRYING 8:11
CYCLE COMPLETE 38:11

Digital Signature #
2BDD77EFE82C174A
