

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2375

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2022 08:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.5 F 0.0  
2:00 119.2 F 1.1  
4:00 138.6 F 3.8  
6:00 167.8 F 7.6  
8:00 186.8 F 12.7  
10:00 218.6 F 5.5  
12:00 228.7 F 10.0  
14:00 236.0 F 8.4  
16:00 240.8 F 11.5  
18:00 237.4 F 9.2  
20:00 251.0 F 15.7  
22:00 262.5 F 22.4  
24:00 272.5 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.3  
1:00 272.7 F 29.0  
2:00 271.7 F 28.8  
3:00 271.8 F 28.7  
4:00 272.4 F 29.0

Min 271.0 F 28.2  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
5:00 190.8 F 0.0  
10:00 198.9 F 0.0  
15:00 193.9 F 0.0  
20:00 191.9 F 0.0  
25:00 199.2 F 0.0  
30:00 205.4 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 24:19  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:02:43  
CYCLE COMPLETE

09/27/2022 09:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_