
Midmark M11 - v1.0.0
Total Cycles: 3630

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2023 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 184.5 F 3.7
4:00 193.8 F 8.4
6:00 213.0 F 11.4
8:00 226.9 F 8.7
10:00 236.7 F 11.1
12:00 239.8 F 11.4
14:00 237.9 F 9.8
16:00 251.5 F 16.3
18:00 263.0 F 22.9
20:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.6 F 29.2
2:00 271.5 F 28.5
3:00 271.8 F 29.0
4:00 272.4 F 29.0

Min 271.1 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 211.2 F 0.0
10:00 215.6 F 0.0
15:00 222.9 F 0.0
20:00 224.8 F 0.0
25:00 224.6 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

08/09/2023 05:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____