
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000948
13:20 09/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:23
STERILIZING 3:06
136.0 C 220kPa 3:06
136.4 C 225kPa 3:36
136.4 C 223kPa 4:06
136.2 C 223kPa 4:36
136.4 C 224kPa 5:06
136.4 C 225kPa 5:36
136.4 C 224kPa 6:06
136.4 C 224kPa 6:36
136.4 C 224kPa 6:36
VENTING 6:36
AIR DRYING 7:26
CYCLE COMPLETE 9:40

Digital Signature #
D00F3EBC7E7A5C57
