

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3219

Name: RITTER2  
Sterilizer ID: V2313228  
12/14/2023 04:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.6 F 0.0  
2:00 181.7 F 3.9  
4:00 203.4 F 9.3  
6:00 220.9 F 7.2  
8:00 234.6 F 10.4  
10:00 233.5 F 7.9  
12:00 242.0 F 11.3  
14:00 247.1 F 13.8  
16:00 259.7 F 20.8  
18:00 270.8 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 272.6 F 29.3  
2:00 271.6 F 28.5  
3:00 272.2 F 29.3  
4:00 272.4 F 29.0  
  
Min 271.1 F 28.4  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
5:00 214.1 F 0.0  
10:00 194.9 F 0.0  
15:00 189.7 F 0.0  
20:00 185.1 F 0.0  
25:00 183.7 F 0.0  
30:00 187.6 F 0.0  
  
Drying Complete

FILLING: 1:57  
HEATING: 18:48  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:03  
CYCLE COMPLETE

12/14/2023 05:39 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_