
Midmark M11 - v1.0.5
Total Cycles: 803

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 01:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.3 F 0.0
2:00 157.1 F 2.8
4:00 170.6 F 6.8
6:00 196.4 F 12.0
8:00 218.9 F 6.7
10:00 233.8 F 12.2
12:00 236.6 F 9.7
14:00 237.1 F 8.9
16:00 249.1 F 15.1
18:00 261.4 F 22.0
20:00 272.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.3
1:00 271.9 F 29.3
2:00 271.9 F 29.3
3:00 272.0 F 29.3
4:00 272.0 F 29.2

Min 271.8 F 29.2
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 227.2 F 0.0
10:00 222.6 F 0.0
15:00 192.9 F 0.0
20:00 202.3 F 0.0
25:00 215.6 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

02/15/2022 02:48 PM
ACCEPTED: X
REJECTED:
Operator: U02