
Midmark M11 - v1.0.5
Total Cycles: 2456

Name: RITTER2
Sterilizer ID: V2313228
06/02/2023 08:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 75.4 F 0.0
4:00 95.0 F 0.0
6:00 114.2 F 0.0
8:00 148.3 F 0.7
10:00 172.7 F 3.9
12:00 203.8 F 7.3
14:00 212.0 F 11.0
16:00 215.3 F 5.8
18:00 226.9 F 7.5
20:00 236.4 F 11.5
22:00 233.5 F 8.1
24:00 245.1 F 13.5
26:00 243.8 F 12.3
28:00 255.7 F 18.5
30:00 266.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 29.0
2:00 272.2 F 29.2
3:00 272.0 F 28.7
4:00 272.7 F 29.4

Min 271.6 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 196.1 F 0.0
10:00 190.1 F 0.0
15:00 193.9 F 0.0
20:00 196.1 F 0.0
25:00 194.9 F 0.0
30:00 194.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:37
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:09:35
CYCLE COMPLETE

06/02/2023 09:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____