
Midmark M11 - v1.0.0
Total Cycles: 878

Name: RITTER 1
Sterilizer ID: V2178198
06/09/2021 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.8 F 0.0
2:00 133.5 F 0.7
4:00 143.8 F 2.0
6:00 175.5 F 4.6
8:00 192.1 F 7.4
10:00 206.0 F 10.8
12:00 219.1 F 6.4
14:00 227.8 F 7.9
16:00 236.6 F 12.1
18:00 235.5 F 8.7
20:00 244.0 F 12.3
22:00 244.9 F 12.4
24:00 255.3 F 18.0
26:00 265.1 F 24.0
28:00 271.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 29.1
2:00 272.4 F 28.7
3:00 271.8 F 28.3
4:00 271.7 F 28.5

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 227.1 F 0.0
10:00 231.0 F 0.0
15:00 230.6 F 0.0
20:00 229.6 F 0.0
25:00 217.3 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:04
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:06:46
CYCLE COMPLETE

06/09/2021 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____