
Midmark M11 - v1.0.0
Total Cycles: 4635

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2024 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.7 F 0.0
2:00 164.5 F 1.8
4:00 169.2 F 5.0
6:00 192.4 F 8.7
8:00 212.0 F 11.4
10:00 220.1 F 7.0
12:00 231.1 F 11.4
14:00 230.8 F 7.5
16:00 243.4 F 13.0
18:00 241.4 F 11.2
20:00 253.1 F 17.1
22:00 263.5 F 23.1
24:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.0
2:00 271.3 F 28.3
3:00 272.6 F 29.1
4:00 271.7 F 28.5

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 216.6 F 0.0
10:00 220.8 F 0.0
15:00 230.1 F 0.0
20:00 238.5 F 0.0
25:00 228.5 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 24:17
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:17
CYCLE COMPLETE

06/13/2024 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____