
Midmark M11 - v1.0.0
Total Cycles: 3243

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2023 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.7 F 0.0
2:00 193.4 F 3.3
4:00 206.0 F 8.3
6:00 222.4 F 9.7
8:00 232.8 F 9.9
10:00 236.8 F 8.4
12:00 243.5 F 12.0
14:00 246.0 F 13.2
16:00 259.0 F 20.3
18:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.2 F 28.8
2:00 271.5 F 28.4
3:00 272.3 F 28.9
4:00 271.5 F 28.2

Min 271.3 F 28.2
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 234.4 F 0.0
10:00 228.7 F 0.0
15:00 228.5 F 0.0
20:00 230.4 F 0.0
25:00 230.2 F 0.0
30:00 205.4 F 0.0

Drying Complete

FILLING: 2:22
HEATING: 18:47
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

05/16/2023 02:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____