
Midmark M11 - v1.0.0
Total Cycles: 3193

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.6 F 0.0
2:00 190.5 F 0.8
4:00 205.5 F 5.2
6:00 217.8 F 10.9
8:00 224.1 F 6.3
10:00 237.5 F 12.0
12:00 237.2 F 9.9
14:00 237.5 F 8.5
16:00 250.7 F 15.8
18:00 263.1 F 22.9
20:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.1
2:00 271.6 F 28.4
3:00 272.6 F 29.3
4:00 272.0 F 28.7

Min 271.5 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 211.8 F 0.0
10:00 232.2 F 0.0
15:00 231.2 F 0.0
20:00 224.4 F 0.0
25:00 222.3 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

05/03/2023 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____