
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001424
13:21 14/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:59
STERILIZING 4:22
136.4 C 220kPa 4:22
136.4 C 224kPa 4:52
136.1 C 222kPa 5:22
136.4 C 225kPa 5:52
136.4 C 221kPa 6:22
136.4 C 225kPa 6:52
136.0 C 220kPa 7:22
136.4 C 226kPa 7:52
136.4 C 226kPa 7:52
VENTING 7:52
AIR DRYING 8:47
CYCLE COMPLETE 23:47

Digital Signature #
C0A7EF4E74D1D299
