
Midmark M11 - v1.0.5
Total Cycles: 2604

Name: RITTER2
Sterilizer ID: V2313228
07/04/2023 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 181.3 F 3.1
4:00 186.3 F 7.7
6:00 215.0 F 13.3
8:00 223.3 F 8.8
10:00 233.2 F 10.2
12:00 240.3 F 12.2
14:00 240.3 F 11.2
16:00 254.1 F 18.2
18:00 266.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.5 F 29.9
2:00 271.4 F 29.1
3:00 272.7 F 30.0
4:00 272.2 F 29.5

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 222.8 F 0.0
10:00 234.2 F 0.0
15:00 225.4 F 0.0
20:00 223.0 F 0.0
25:00 224.0 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

07/04/2023 03:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____