

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000950  
18:04 23/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:40  
PRESSURIZING 4:41  
STERILIZING 5:58  
136.3 C 231kPa 5:58  
136.4 C 233kPa 6:28  
136.4 C 233kPa 6:58  
136.3 C 233kPa 7:28  
136.2 C 231kPa 7:58  
136.0 C 231kPa 8:28  
136.2 C 232kPa 8:58  
136.4 C 233kPa 9:28  
136.4 C 233kPa 9:28  
VENTING 9:29  
AIR DRYING 10:29  
CYCLE COMPLETE 25:29  
-----

Digital Signature #  
85AF03D57B1EE869  
-----