
Midmark M11 - v1.0.0
Total Cycles: 3782

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.2 F 0.0
2:00 121.0 F 0.0
4:00 145.9 F 0.5
6:00 174.0 F 2.7
8:00 190.9 F 5.6
10:00 217.4 F 9.5
12:00 228.5 F 13.5
14:00 229.5 F 7.7
16:00 239.5 F 12.0
18:00 236.3 F 8.8
20:00 242.4 F 11.3
22:00 245.5 F 12.7
24:00 256.8 F 18.6
26:00 266.8 F 24.8
28:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.1 F 28.5
2:00 272.6 F 28.8
3:00 272.0 F 28.8
4:00 272.1 F 28.4

Min 271.6 F 28.3
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 243.7 F 0.0
10:00 230.9 F 0.0
15:00 227.6 F 0.0
20:00 217.0 F 0.0
25:00 209.5 F 0.0
30:00 184.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 28:15
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:06:45
CYCLE COMPLETE

09/15/2023 02:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____