
Midmark M11 - v1.0.0
Total Cycles: 2195

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2022 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 181.3 F 4.3
4:00 194.2 F 9.0
6:00 208.9 F 9.3
8:00 228.0 F 8.2
10:00 236.9 F 13.3
12:00 237.7 F 10.3
14:00 235.9 F 8.1
16:00 249.0 F 14.7
18:00 260.3 F 21.0
20:00 270.5 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.0 F 28.9
2:00 272.1 F 28.7
3:00 271.3 F 28.4
4:00 272.6 F 29.2

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 236.1 F 0.0
10:00 225.8 F 0.0
15:00 230.2 F 0.0
20:00 229.0 F 0.0
25:00 230.7 F 0.0
30:00 232.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

08/03/2022 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____