
Midmark M11 - v1.0.5
Total Cycles: 938

Name: RITTER2
Sterilizer ID: V2313228
03/31/2022 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.0 F 0.0
2:00 205.4 F 3.9
4:00 214.0 F 10.0
6:00 224.7 F 7.1
8:00 238.6 F 12.9
10:00 240.3 F 11.9
12:00 241.2 F 11.7
14:00 255.2 F 18.9
16:00 267.3 F 26.3
18:00 272.3 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.7
1:00 272.4 F 29.7
2:00 272.8 F 30.0
3:00 272.6 F 30.3
4:00 272.0 F 29.7

Min 271.9 F 29.6
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 220.9 F 0.0
10:00 232.9 F 0.0
15:00 227.2 F 0.0
20:00 228.9 F 0.0
25:00 228.1 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 18:02
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

03/31/2022 01:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____