
Midmark M11 - v1.0.0
Total Cycles: 1508

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2022 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.0 F 0.0
2:00 156.0 F 1.7
4:00 178.9 F 4.3
6:00 195.9 F 7.8
8:00 209.9 F 12.0
10:00 218.4 F 4.3
12:00 220.7 F 8.4
14:00 231.3 F 12.6
16:00 233.2 F 8.5
18:00 243.8 F 13.1
20:00 242.1 F 11.3
22:00 253.7 F 17.2
24:00 264.3 F 23.4
26:00 271.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.5 F 28.7
2:00 271.9 F 28.7
3:00 272.1 F 28.5
4:00 272.4 F 29.0

Min 271.5 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.3 F 0.0
5:00 235.7 F 0.0
10:00 226.8 F 0.0
15:00 227.5 F 0.0
20:00 227.2 F 0.0
25:00 231.5 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:04:12
CYCLE COMPLETE

01/11/2022 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____