
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001821
8:32 10/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 2:01
STERILIZING 3:03
136.5 C 225kPa 3:03
136.4 C 226kPa 3:33
136.1 C 223kPa 4:03
136.2 C 223kPa 4:33
136.4 C 226kPa 5:03
136.4 C 226kPa 5:33
136.4 C 226kPa 6:03
136.4 C 226kPa 6:33
136.4 C 226kPa 6:33
VENTING 6:33
AIR DRYING 7:22
CYCLE COMPLETE 7:42

Digital Signature #
3AF526EEED752E0B
