
Midmark M11 - v1.0.0
Total Cycles: 2150

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.7 F 0.0
2:00 165.5 F 1.9
4:00 184.1 F 6.5
6:00 207.2 F 11.2
8:00 221.5 F 5.3
10:00 236.3 F 11.0
12:00 233.0 F 8.0
14:00 241.4 F 10.9
16:00 245.5 F 13.1
18:00 258.1 F 19.8
20:00 269.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.6 F 29.1
2:00 271.5 F 28.3
3:00 272.4 F 29.2
4:00 272.3 F 28.8

Min 271.2 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 220.2 F 0.0
10:00 228.4 F 0.0
15:00 231.9 F 0.0
20:00 222.4 F 0.0
25:00 226.7 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

07/19/2022 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____