

-----  
Midmark M11 - v1.0.0  
Total Cycles: 729

Name: RITTER 1  
Sterilizer ID: V2178198  
03/18/2021 08:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.5 F 0.0  
2:00 147.1 F 1.2  
4:00 159.4 F 4.2  
6:00 188.9 F 8.4  
8:00 206.6 F 13.5  
10:00 218.8 F 6.5  
12:00 232.3 F 11.7  
14:00 232.9 F 8.3  
16:00 241.0 F 10.8  
18:00 244.9 F 12.8  
20:00 257.7 F 19.6  
22:00 269.0 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.8 F 29.2  
2:00 272.0 F 28.5  
3:00 271.2 F 28.2  
4:00 272.4 F 29.2

Min 271.2 F 28.1  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.1 F 0.0  
5:00 219.9 F 0.0  
10:00 221.6 F 0.0  
15:00 229.3 F 0.0  
20:00 240.7 F 0.0  
25:00 220.9 F 0.0  
30:00 217.9 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 23:08  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:01:20  
CYCLE COMPLETE

03/18/2021 09:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_