
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000179
7:31 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:53
STERILIZING 5:24
136.0 C 220kPa 5:24
136.3 C 224kPa 5:54
136.4 C 222kPa 6:24
136.4 C 224kPa 6:54
136.2 C 223kPa 7:24
136.4 C 225kPa 7:54
136.4 C 225kPa 8:24
136.3 C 224kPa 8:54
136.3 C 224kPa 8:54
VENTING 8:54
AIR DRYING 9:57
CYCLE COMPLETE 24:57

Digital Signature #
8BE91896B51625DA
