
Midmark M11 - v1.0.0
Total Cycles: 2286

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2022 03:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.6 F 0.0
2:00 183.3 F 1.8
4:00 206.5 F 7.2
6:00 223.6 F 12.4
8:00 230.1 F 8.3
10:00 241.1 F 11.5
12:00 242.5 F 11.9
14:00 241.0 F 10.7
16:00 254.4 F 17.6
18:00 266.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.8
2:00 271.5 F 28.0
3:00 272.5 F 29.0
4:00 272.1 F 28.4

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 238.1 F 0.0
10:00 234.4 F 0.0
15:00 231.1 F 0.0
20:00 227.9 F 0.0
25:00 228.9 F 0.0
30:00 230.5 F 0.0

Drying Complete

FILLING: 1:49
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

09/02/2022 04:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____