
Midmark M11 - v1.0.0
Total Cycles: 5217

Name: RITTER 1
Sterilizer ID: V2178198
01/02/2025 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.2 F 0.0
2:00 136.6 F 0.0
4:00 162.5 F 0.6
6:00 184.0 F 3.9
8:00 201.0 F 7.5
10:00 214.1 F 11.6
12:00 220.7 F 5.4
14:00 232.1 F 9.9
16:00 235.9 F 9.5
18:00 242.2 F 12.1
20:00 240.0 F 10.5
22:00 252.5 F 16.6
24:00 263.4 F 23.0
26:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.8 F 29.1
2:00 271.8 F 28.3
3:00 272.4 F 29.1
4:00 272.3 F 28.8

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 193.9 F 0.0
10:00 191.1 F 0.0
15:00 195.7 F 0.0
20:00 190.6 F 0.0
25:00 190.4 F 0.0
30:00 189.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

01/02/2025 02:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____