
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000471
8:15 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:44
PRESSURIZING 0:51
STERILIZING 1:52
136.4 C 224kPa 1:52
136.3 C 224kPa 2:22
136.3 C 223kPa 2:52
136.4 C 224kPa 3:22
136.3 C 223kPa 3:52
136.4 C 223kPa 4:22
136.0 C 219kPa 4:52
136.4 C 223kPa 5:22
136.4 C 223kPa 5:22
VENTING 5:22
AIR DRYING 6:15
CYCLE COMPLETE 6:26

Digital Signature #
159E4C3CCEE1187D
